SANA

Certificate

This is to certify that

The laboratory SANA Srl accredited by ACCREDIA with № 0747
according to the UNI EN ISO 17025:2018 has performed

a virucidal activity test (MERS-SARS)
using Human Coronavirus 229E ATCC® VR-740/VR-19 strain from 10th to 31th July 2020

with STEAM GENERATOR "EV" of the company E V INTERNATIONAL Ltd.
together with the product EV H_2O_2 SILVER BIO and the product EV H_2O_2 SILVER SPRAY

in accordance with UNI EN 16777:2018
Chemical Disinfectants and antiseptics – Quantitative non-porous surface test without mechanical action for the evaluation of virucidal activity of chemical disinfectants used in the medical area – Test method and requirements (phase 2/step 2).

In particular, the results obtained by the test show the reduction of the virus viability by more than 4 log by spraying the product for 4-5 seconds on the test surface (contact time: 30 seconds) i.e. the viral titre was not detected.

Dr. Marta Colatosti
Laboratory director

SANA s.r.l. FROSINONE

signature and stamp