# CERTIFICATE

*This is to certify that*

*The laboratory SANA Srl accredited by ACCREDIA with № 0747*
*according to the UNI EN ISO 17025:2018 has performed*

*a test (AFNOR - NF T72-110)*
*using bactericidal activity, including Staphylococcus aureus ATCC 6538*
*from 5th to 12th October 2023*

*with STEAM GENERATOR "EV" of the company EV INTERNATIONAL Ltd.*
*in accordance with AFNOR - NF T72-110 STANDARD for Steam Disinfection Device (SDD)*

*Quantitative surface test with mechanical action for the evaluation*
*of bactericidal activity of dry steam used in the medical area.*

*This certificate attests to the efficacy of the Steam Disinfection Device under specified procedure.*
*It confirms the efficacy result measured > 99,999% in clean conditions of the viability of the counts*
*under examination for bactericidal activity, endorsing its suitability for disinfection purposes.*

SANA

*Dr. Marta Colatosti*
*Laboratory director*

SANA s.r.l. FROSINONE

signature and stamp